

Technical Datasheet

Article No. 11111

pH Calibration kit 4.01 -7.01 -10.01

For laboratory use.

Parameter	Value
Density	1.01 – 1.02 g/ml
pH (20 °C)	4.00 – 4.02
pH (20 °C)	7.00 – 7.02
pH (20 °C)	10.00 – 10.02